



Tilt signals

VP5  
 VCC 3012 U3 470 J7  
 GND Temp R34 CON1  
 25° ==> 375 mV  
 60° ==> 550 mV

Out = 5\*(.009\*Pressure\_kPa - .095)  
 if P >15 kPa

Temperature tests, 26-jul-2013  
 On optical bench, wrapped into rags, after 3 hours:  
 Current capteur\_TSIC Totor LT1028 50 ohms

# QUADVIDE pour Advanced Virgo

Title		
Size	Number	Revision
A4		
Date:	14/10/2016	Sheet of
File:	C:\schemas\quadrantvide\quadvide.sch	Drawn By: